

Advanced Hemostatic and Wound Healing Dressing for Donor Sites

100% natural plant-based

Instructions for Medical Professionals

Indications for Use

Emergency and therapeutic use in the control of bleeding from the skin and other surface wounds where temporary control of bleeding is required.

Directions

1. Apply DonorSeal Matrix to the donor site; it will turn into a gel upon contact with blood and exudate to seal the wound. Use additional layers of DonorSeal as needed.

2. Place a non-adherent secondary dressing over DonorSeal to keep moist.

Wrap sterile gauze over the non-adherent dressing, followed by an ace bandage for compression. Change the outer gauze daily or as needed for drainage.

4. Remove all dressings, gently clean the site with sterile water or sterile saline, being careful not to disrupt the wound healing.

WARNINGS:

Longer compression times may be required for patients who are obese, hypertensive, or are on anticoagulants.

When removing DonorSeal and/or any secondary dressing, be careful not to disturb the clot.

Never peel DonorSeal from a wound, as re-bleeding could occur. If DonorSeal dries out, soak thoroughly to rehydrate before removal.

In case of swelling or redness, discontinue use.

Do not reuse, as device will no longer be sterile after initial use, and could result in infection


Cover DonorSeal with a wrap or adhesive bandage to maintain pressure if the patient is to be moved.

STERILE: Sterility guaranteed unless individual wrapper is opened or damaged.

STORAGE: For maximum shelf life, storage temperature of 15-30 °C (59-86 °F) is recommended.



Not made with natural rubber latex





اس

Manufacturer: LifeScience PLUS, Inc. P.O. Box 60783, Palo Alto, CA 94306 2520A Wyandotte St., Mountain View, Toll Free: 1-877-587-5433 - U.S.A. 1-650-656-8172 - International Fax: 1-650-336-1130 www.LifeSciencePlus.com Email: info@lifescienceplus.com Manufactured in the USA View, CA 94043

LifeScience PLUS

© 2024 LifeScience PLUS, Inc. All Rights Reserved.